

	EIF	MON	GON
Calibration sequence	2 h every 3 months	6 h every 2 weeks	6 h every 2 weeks
Target sequence	30 min every 2 weeks	30 min every 12 h	30 min every 12 h
Accuracy (ppm)	0.13	-0.04	-0.07
Repeatability (ppm)	0.38	0.10	0.07