

Brewer		Correlation, median of differences, SD ( $1\sigma$ ) of differences				
(Mk)	obs.	306.3 nm	310.1 nm	313.5 nm	316.8 nm	320.1 nm
Calibrated by the Langley method						
183 (III)	4695	0.934, 0.0002, 0.0105	0.967, 0.0031, 0.0073	0.975, 0.0033, 0.0064	0.981, 0.0047, 0.0055	0.985, 0.0039, 0.0050
Calibrated by transfer from Brewer #185						
070 (IV)	438	0.783, -0.0003, 0.0365	0.948, 0.0000, 0.0145	0.955, 0.0000, 0.0136	0.955, -0.0001, 0.0133	0.955, -0.0001, 0.0133
075 (IV)	303	0.863, -0.0007, 0.0288	0.972, -0.0002, 0.0133	0.972, 0.0000, 0.0134	0.978, -0.0001, 0.0115	0.976, 0.0000, 0.0119
186 (III)	509	0.931, 0.0001, 0.0127	0.960, 0.0001, 0.0092	0.967, 0.0001, 0.0083	0.971, 0.0001, 0.0078	0.973, 0.0000, 0.0075
201 (III)	407	0.907, 0.0001, 0.0106	0.946, 0.0003, 0.0074	0.949, 0.0002, 0.0069	0.956, 0.0001, 0.0064	0.955, 0.0001, 0.0063
202 (III)	464	0.983, 0.0002, 0.0090	0.992, 0.0000, 0.0062	0.993, 0.0001, 0.0057	0.994, 0.0001, 0.0054	0.994, 0.0000, 0.0053
Median	438	0.907, 0.0001, 0.0127	0.960, 0.0000, 0.0092	0.967, 0.0001, 0.0083	0.971, 0.0001, 0.0078	0.973, 0.0000, 0.0075