

Interactive comment on “Prompt deliquescence and efflorescence of aerosol nanoparticles” by G. Biskos et al.

M. Cubison

michael.cubison@colorado.edu

Received and published: 13 September 2006

Further to comment 4) of Hämeri and Laaksonen, it is noted that the University of Manchester constructed an HTDMA with fast RH response which was described by Cubison et al (2005); the modified HTDMA instrument included automatic valves which could switch out the sheath humidification system after a large step change in RH, such that an equilibrium could be achieved between the aerosol and sheath (ie, DMA) RH values. Would the authors please consider referencing this study and following the suggestion in the previous comment of this thread that a discussion of the possible effects of any RH discrepancy be included in the revised paper.

A modified hygroscopic tandem DMA and a data retrieval method based on optimal estimation. M.J.Cubison, M.Gysel and H.Coe, J Aero.Sci., 36(7), p846-865 (2005)

Interactive comment on Atmos. Chem. Phys. Discuss., 6, 7051, 2006.

ACPD

6, S3119–S3120, 2006

Interactive
Comment

Full Screen / Esc

Printer-friendly Version

Interactive Discussion

Discussion Paper

S3120

EGU