

Interactive
Comment

Interactive comment on “High levels of reactive gaseous mercury observed at a high elevation research laboratory in the Rocky Mountains” by X. Faïn et al.

S. Lyman

slyman@uw.edu

Received and published: 4 September 2009

This manuscript states that the Tekran 2537 mass flow meter is referenced to a standard temperature and pressure of 20 degrees C and 1 ATM. Tekran instruments actually use 0 degrees C and 1 ATM, an 8% difference.

Interactive comment on Atmos. Chem. Phys. Discuss., 9, 15641, 2009.

Full Screen / Esc

Printer-friendly Version

Interactive Discussion

Discussion Paper

